Vanguard Laboratory
2635 Parkmont Ln
Olympia, Wa 98502
360-967-7010

Certificate of Analysis

Glutathione 1500 mg

Report To:
Solyn LLC

Compound: Glutathione
Amount: 1500 mg
Laboratory ID: V260508-9 002
Date Reported: 6/19/2026
Lot Number: 1248



Chromatogram for vial A shown above.

Analysis	Method	Result (per Vial)		
Chromatographic Purity	HPLC-UV/VIS	99.15%	99.56%	99.45%
Assay	HPLC-UV/VIS	1477 mg	1457 mg	1485 mg



Report By: Dustin Newman, Laboratory Director on 6/19/2026

Approved By: Tori Johnson, Operations Manager on 6/19/2026

ISO 17025:2017 ACCREDITED LABORATORY

Please consult A2LA Certificate #6377.01.01 for a list of accredited tests. Samples were received in acceptable condition. The result(s) in this report relate only to the portion of the sample(s) tested. All analyses were performed consistent with the Vanguard Laboratory Quality Management System. Vanguard Laboratory and its staff did not observe or participate in the sample selection process, and cannot confirm the authenticity of the sample or its representativeness of the associated lot/batch.

Vanguard Laboratory
2635 Parkmont Ln
Olympia, Wa 98502
360-967-7010

Certificate of Analysis

Glutathione 1500 mg

Report To:
Solyn LLC

Compound: Glutathione
Amount: 1500 mg
Laboratory ID: V260508-9 002
Date Reported: 6/19/2026
Lot Number: 1248

Analyte	Method	LOQ (ppm)	Result (per Vial)	
Chromium (Cr)	ICP-MS	0.05	ND	ND
Arsenic (As)	ICP-MS	0.01	0.01	ND
Cadmium (Cd)	ICP-MS	0.01	ND	ND
Lead (Pb)	ICP-MS	0.02	ND	ND
Mercury (Hg)	ICP-MS	0.005	ND	ND

Run ID: 260508

Analysis	Method	Result (per Vial)	
Endotoxins	LAL	**A**	Pass - <5.00 EU/mg*
		B	Pass - <5.00 EU/mg*

*Pass/Fail criteria based on the USP/FDA threshold of 5 EU/kg (350 EU total for a 70 kg adult)

Analysis	Method	Result (per Vial)	
Sterility	USP <71>	**A**	Pass - No Growth
		B	Pass - No Growth
		C	Pass - No Growth
		D	Pass - No Growth

Report By: Dustin Newman, Laboratory Director on 6/19/2026

Approved By: Tori Johnson, Operations Manager on 6/19/2026

ND: Non-Detect
LOD: Limit of Quantification

ISO 17025:2017 ACCREDITED LABORATORY

Please consult A2LA Certificate #6377.01.01 for a list of accredited tests. Samples were received in acceptable condition. The result(s) in this report relate only to the portion of the sample(s) tested. All analyses were performed consistent with the Vanguard Laboratory Quality Management System. Vanguard Laboratory and its staff did not observe or participate in the sample selection process, and cannot confirm the authenticity of the sample or its representativeness of the associated lot/batch.

Vanguard Laboratory
2635 Parkmont Ln
Olympia, Wa 98502
360-967-7010

Certificate of Analysis

Glutathione 1500 mg

Report To:
Solyn LLC

Compound: Glutathione
Amount: 1500 mg
Laboratory ID: V260508-9 002
Date Reported: 6/19/2026
Lot Number: 1248

Residual Solvents Analysis			
Analyte	**Method**	**LOQ (ppm)**	**Result (ppm)**
Acetone	GC-MS	3.7	ND
Benzene	GC-MS	0.02	ND
Butanes	GC-MS	3.7	ND
Chloroform	GC-MS	0.02	ND
Cyclohexane	GC-MS	2.9	ND
Dichloromethane	GC-MS	0.45	ND
Ethanol	GC-MS	3.7	ND
Ethyl Acetate	GC-MS	3.7	ND
Heptane	GC-MS	3.7	ND
Hexanes	GC-MS	0.21	ND
Isopropanol	GC-MS	3.7	ND
Methanol	GC-MS	2.2	ND
Pentanes	GC-MS	3.7	ND
Propane	GC-MS	3.7	ND
Toluene	GC-MS	0.66	ND
Total Xylenes	GC-MS	1.6	ND

Analysis	Method	Result (per Vial)	
Container Closure Integrity	Dye Ingress Method	**A**	Pass
		B	Pass
Solubility Analysis	Fully Soluble in 2 mL Water	**A**	Pass
Trifluoroacetic Acid (TFA)	Ion Chromotography	ND - <0.10%	

ND: Non-Detect
LOD: Limit of Quantification

ISO 17025:2017 ACCREDITED LABORATORY

Report By: Dustin Newman, Laboratory Director on 6/19/2026

Approved By: Tori Johnson, Operations Manager on 6/19/2026

Please consult A2LA Certificate #6377.01.01 for a list of accredited tests. Samples were received in acceptable condition. The result(s) in this report relate only to the portion of the sample(s) tested. All analyses were performed consistent with the Vanguard Laboratory Quality Management System. Vanguard Laboratory and its staff did not observe or participate in the sample selection process, and cannot confirm the authenticity of the sample or its representativeness of the associated lot/batch.